

Neptune® Gold & Neptune Bronze

**PLEASE REFER TO THE
RECALL NOTIFICATION LETTER
FOR COMPLETE DETAILS**

Required Actions



0700-001-000 Neptune Gold



0700-007-000 Neptune Bronze

1	Ensure all users of the Neptune device, including surgeons, surgical residents and health profession students or OR assignments, are adequately trained, and are aware of the risks associated with the device as detailed in the device recall notification.	<input type="checkbox"/>
2	It is recommended that all facilities keep a master list of all personnel that have been trained on the use of the Neptune. This list should include all users, such as surgeons, anesthesiologists, residents, nurses and technicians.	<input type="checkbox"/>
3	Inform all users of the Neptune device that additional adverse events have been reported.	<input type="checkbox"/>
4	Ensure warning labels, previously supplied by Stryker, are present on all Neptune device(s). If you require additional labels, please contact Stryker at 855-458-7441 or 269-389-2316.	<input type="checkbox"/>
5	Complete the business reply form, acknowledging you completed the above actions.	<input type="checkbox"/>